


PATIENT INSTRUCTIONS

Kinex KS2™ Shoulder CPM

To change Abduction and External Rotation:

1. Press the **STOP** button.
2. To change abduction, press the **MENU** button once; M1 should appear on the screen.
3. Press the upper left **Abduction** button on hand controller.
4. Press **(+)** to increase and **(-)** to decrease.
5. To change External Rotation, press the upper right **External Rotation** button on hand controller.
6. Press **(+)** to increase and **(-)** to decrease.
7. Press the **START** button to save.
8. Press the **START** button to begin therapy.



Distributed by:

KINEX
Phase I Rehab™